

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Switch mode power supply

Name and address of the applicant

TDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE
EX34 8ES UNITED KINGDOM

Name and address of the manufacturer

TDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE
EX34 8ES UNITED KINGDOM

Name and address of the factory

Note: When more than one factory, please report on page 2

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE EX34 8ES
UNITED KINGDOM Additional Information on page 2

Ratings and principal characteristics

Input: 100-240Vac; 47-63Hz; 2.2Arms Max.
Output: (see test report for details)
Class I

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

CUS150M
See Page 2

Additional information (if necessary may also be reported on page 2)

National Differences specified in the CB Test Report

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.2), IEC 60601-1(ed.2);am1, IEC 60601-1(ed.2);am2

As shown in the Test Report Ref. No. which forms part of this Certificate

E349607-A44-CB-1 issued on 2017-09-25

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-09-25

Original Issue Date: 2017-04-28

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-63104-A1-UL

Model Details:

CUS150M Unit Product Code:
CUS150M-xxVx/y

Where:

-xxVx = Channel 1 output voltage from within the output voltage adjustment range from the Output Parameters Table below.

Where:

y = Unit options from list of standard unit options below, or non-safety related model differences.

/M = Molex connectors

/E = Single fuse in the live line

/U = U chassis

/A = Cover and U chassis

/F = Top fan, cover and U chassis

Factories:

PANYU TRIO MICROTRONIC CO LTD
SHIJI INDUSTRIAL ESTATEDONGYONG NANSHA GUANGZHOU GUANGDONG
CHINA

Additional Information:

The original report was modified to include the following changes/additions: leakage current test repeated per the client's request, new measurements replaced the previous test results.

Additional information (if necessary)



UL (US), 333 Pflingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-09-25

Original Issue Date: 2017-04-28

Signature:

Jan-Erik Storgaard