



TDK-Lambda Americas Inc
405 Essex Road
Neptune NJ 07753 USA
Tel: +1 732 922 9300
Fax: +1 732 922 9334
www.us.tdk-lambda.com/hp

DECLARATION OF CONFORMITY GEN 3U AND QS SERIES

We, TDK-Lambda Americas Inc., of 405 Essex Road, Neptune, NJ 07753, USA declare under our sole responsibility that the GEN3U and QS series as detailed on the attached products covered sheet comply with the provisions of the following European Directives and are eligible to bear the CE mark:


Low Voltage Directive	Directive 2014/35/EU
EMC Directive	Directive 2014/30/EU
RoHS 2	Directive 2011/65/EU

Assurance of conformance of the described product with the provisions of the stated EC Directives is given through compliance to the following standards:

Electrical Safety:	EN60950-1:2006 + A2:2013
Electromagnetic Emissions:	EN 61326-1:2013

These products are high-power equipment, with input power >1 kW, for professional use and installation, and carries the CE mark accordingly. These products are for use in Class A, ITE environment only, as defined by EN 61326-1:2013 and EN 61204-3:2000.

Our European Representative in the EU is TDK-Lambda UK Limited, located at Kingsley Avenue, Ilfracombe, Devon, EX34 8ES UK. Further all products covered by this declaration are manufactured in accordance with ISO9001:2008 which ensure continued compliance of the products with the requirements of the Low Voltage and the EMC directives.

Name of Authorized Signatory	Martin Southam
Signature of Authorized Signatory	
Position of Authorized Signatory	Marketing Director, TDK-Lambda EMEA
Date	1 st August 2016
Date series first CE marked	1 st October 2008
Place where signed	Ilfracombe, Devon, England

PRODUCTS COVERED SHEET FOR THE GEN3U AND QS SERIES

Product Name: GEN 3U Series Switch Mode DC Power Supply
QS Series Switch Mode DC Power Supply

Models: GEN AAA-BBB-KKK-Z
QS AAA-BBB-KKK-Z

A= Volts (7.5V to 600V)

B= Current (25A to 1000A depending on power not to exceed 15kW).

KKK = alphanumeric designation for options that do NOT affect EMC or Safety

Z= 3P208 for 208Vac 3 phase input or 3P400 for 400Vac 3 phase input.

Z may be followed by a 4 digit special number which covers minor differences not affecting safety or EMC.