

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product	Switch mode power supplies
Name and address of the applicant	TDK-Lambda UK Ltd Kingsley Avenue, Ilfracombe Devon, EX34 8ES UNITED KINGDOM
Name and address of the manufacturer	TDK-Lambda UK Ltd Kingsley Avenue, Ilfracombe Devon, EX34 8ES UNITED KINGDOM
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	TDK-Lambda UK Ltd Kingsley Avenue, Ilfracombe Devon, EX34 8ES UNITED KINGDOM <input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	See Page 2
Trademark (if any)	
Type of Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	QM5, QM7, QM8, QS5, QS7, See Page 2
Additional information (if necessary may also be reported on page 2)	<input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1
As shown in the Test Report Ref. No. which forms part of this Certificate	E349607-D1002-2/A0/C0-ULCB issued on 2018-06-01

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2018-06-27

Signature:

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Jan-Erik Storgaard



Ref. Certif. No.

**DK-74224-UL**

**Model Details:**

QM5,QM7,QM8,QS5,QS7 (followed by alphanumeric characters - see Test Report model differences for details of models and nomenclature)

**Factories:**

PANYU TRIO MICROTRONIC CO. LTD  
SHIJI INDUSTRIAL ESTATE DONGYONG, NANSHA, GUANGZHOU GUANGDONG  
CHINA

**Ratings:**

QM5 & QS5 (700W): 100-240Vac nom, 47-63Hz, 11A rms max or 144-318Vdc nom, 7A max.  
QM5 & QS5 (1200W): 200-240Vac nom, 47-63Hz, 9A rms max or 239-318Vdc nom, 7A max.  
QM7 & QS7 (1200W): 100-240Vac nom, 47-63Hz, 19A rms max.  
QM7 & QS7 (1500W): 166.7-240Vac nom, 47-63Hz, 14A rms max.  
QM8 (1200W): 100-240Vac nom, 47-63Hz, 19A rms max.  
QM8 (1500W): 166.7-240Vac nom, 47-63Hz, 14A rms max.

**Additional Information:**

Risk Management was not included in this investigation.

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014. National Differences specified in the CB Test Report.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2018-06-27

Signature:

Jan-Erik Storgaard