


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

|   |   |
|---|---|
| Product   | Medical Switch Mode Power Supply  |
| Name and address of the applicant   | TDK-LAMBDA UK LTD<br>KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES<br>UNITED KINGDOM   |
| Name and address of the manufacturer  | TDK-LAMBDA UK LTD<br>KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES<br>UNITED KINGDOM   |
| Name and address of the factory<br><i>Note: When more than one factory, please report on page 2</i> | TDK-LAMBDA UK LTD<br>KINGSLEY AVE ILFRACOMBE EX34 8ES<br>UNITED KINGDOM<br><input checked="" type="checkbox"/> Additional Information on page 2 |
| Ratings and principal characteristics   | See Page 2<br><br>TDK-Lambda  |
| Trademark (if any)  |    |
| Type of Customer's Testing Facility (CTF) Stage used  |   |
| Model / Type Ref.   | Vega 450, Vega 650, Vega 900, Vega Lite 550, Vega Lite 750<br>See Page 2  |
| Additional information (if necessary may also be reported on page 2)                                | National Differences specified in the CB Test Report.<br><input type="checkbox"/> Additional Information on page 2                              |
| A sample of the product was tested and found to be in conformity with                               | IEC 60601-1(ed.2), IEC 60601-1(ed.2);am1, IEC 60601-1(ed.2);am2   |
| As shown in the Test Report Ref. No. which forms part of this Certificate                           | E349607-A20-CB-2 issued on 2018-01-03   |

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2018-01-17

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-70000-UL**

**Model Details:**

Vega 450, Vega 650, Vega 900, Vega Lite 550, Vega Lite 750 (see Model Difference in test report for details of models and nomenclature)

**Factories:**

PANYU TRIO MICROTRONICS CO LTD  
SHIJI INDUSTRIAL ESTATE DONGYONG NANSHAGUANGZHOU GUANGDONG 511453  
CHINA

**AVNET INC.**

AVNET LOGISTICS 6700 WEST MORELOS PLACE CHANDLER AZ 85226  
UNITED STATES

**Ratings:**

Vega 450 and Vega Lite 550.

PSUs with cooling option F and without xEW and xFW options:

Input voltage: 94.5-240 V ac nom., 85-264V ac max., 47-63 Hz, 8.5 A rms max.

All other PSUs:

Input voltage: 100-240 V ac nom., 90-264V ac max., 47-63 Hz, 8.5 A rms max.

Vega 650, Vega Lite 750 and Vega 900.

PSUs with cooling option F and without xEW and xFW options:

Input voltage: 94.5-240 V ac nom., 85-264V ac max., 47-63 Hz, 12 A rms max.

All other models:

Input voltage: 100-240 V ac nom., 90-264V ac max., 47-63 Hz, 11 A rms max.

(See Model Differences for details of ratings)

**Additional information (if necessary)**



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2018-01-17

Signature:

Jan-Erik Storgaard